



### Désignation

Bloc en bois de fixation d'emboîture.

### Destination

Bloc en bois de fixation d'emboîture entrant dans la réalisation d'une prothèse du membre inférieur.

- ☞ Limite de poids : 75 kg

### Restriction d'utilisation / Précautions d'emploi

Couple de serrage 5 Nm (utiliser la clé dynamométrique (Réf. : ZOC102))

- ☞ Le bloc en bois doit être utilisé suivant l'appréciation de l'orthoprothésiste agréé en fonction du patient à appareiller
- ☞ La maintenance du produit reste à l'appréciation de l'orthoprothésiste selon le cadre de l'utilisation
- ☞ L'assemblage et l'ajustage doivent être effectués par un personnel qualifié

### Entretien / Stockage

- ☞ Stocker le produit à l'abri de l'humidité et dans son emballage d'origine (sur lequel est inscrit son numéro de lot)
- ☞ La fréquence des visites de contrôle de l'appareil reste à l'appréciation de l'orthoprothésiste en fonction du patient à appareiller

### Garantie

La garantie est valable 1 an pour une utilisation conforme aux instructions du fabricant à partir de la date d'achat du produit.

*NB : Pour toutes réclamations veuillez joindre un formulaire de retour dûment complété, tout formulaire incomplet ne pourra être traité.*

### Marquage CE

Groupe G2M déclare, en sa qualité de distributeur, que toutes ses pièces sont conformes à la directive 93/42/CE.



### Désignation

Bloc en bois de fixation d'emboîture.

### Destination

Bloc en bois de fixation d'emboîture entrant dans la réalisation d'une prothèse du membre inférieur.

- ☞ Limite de poids : 75 kg

### Restriction d'utilisation / Précautions d'emploi

Couple de serrage 5 Nm (utiliser la clé dynamométrique (Réf. : ZOC102))

- ☞ Le bloc en bois doit être utilisé suivant l'appréciation de l'orthoprothésiste agréé en fonction du patient à appareiller
- ☞ La maintenance du produit reste à l'appréciation de l'orthoprothésiste selon le cadre de l'utilisation
- ☞ L'assemblage et l'ajustage doivent être effectués par un personnel qualifié

### Entretien / Stockage

- ☞ Stocker le produit à l'abri de l'humidité et dans son emballage d'origine (sur lequel est inscrit son numéro de lot)
- ☞ La fréquence des visites de contrôle de l'appareil reste à l'appréciation de l'orthoprothésiste en fonction du patient à appareiller

### Garantie

La garantie est valable 1 an pour une utilisation conforme aux instructions du fabricant à partir de la date d'achat du produit.

*NB : Pour toutes réclamations veuillez joindre un formulaire de retour dûment complété, tout formulaire incomplet ne pourra être traité.*

### Marquage CE

Groupe G2M déclare, en sa qualité de distributeur, que toutes ses pièces sont conformes à la directive 93/42/CE.



### Designation

Wood block for socket adaptation.

### Purpose

Wood block used in the makin of lower limb prothesis.

- Weight limit : 75 kg

### Restriction on use / Precautions for use

Tightening torques 5 Nm (use key (Ref. : ZOC102))

- The product must be used within the recommended weight limits and in accordance with the assessment made by an approved ortho-prosthetist depending on the patient concerned
- Product maintenance is subject to an assessment made by an approved ortho-prosthetist depending on the conditions of use
- All assembly and adjustment work must be carried out by qualified personnel

### Cleaning / Storage

- Stock the product in a dry place, in its original packing (on which the batch number is marked)
- The frequency of inspections to check the equipment is subject to the assessment made by the ortho-prosthetist depending on the patient concerned

### Guarantee

The guarantee is valid for one year as from the date of purchase of the product, if the product has been used in accordance with the manufacturer's instructions.

*N.B. In the event of a complaint, please enclose a return form correctly filled in, incomplete forms cannot be dealt with.*

### EC markings

Groupe G2M hereby declares, in its quality as the manufacturer, that all its parts comply with directive 93/42/EC.



### Designation

Wood block for socket adaptation.

### Purpose

Wood block used in the makin of lower limb prothesis.

- Weight limit : 75 kg

### Restriction on use / Precautions for use

Tightening torques 5 Nm (use key (Ref. : ZOC102))

- The product must be used within the recommended weight limits and in accordance with the assessment made by an approved ortho-prosthetist depending on the patient concerned
- Product maintenance is subject to an assessment made by an approved ortho-prosthetist depending on the conditions of use
- All assembly and adjustment work must be carried out by qualified personnel

### Cleaning / Storage

- Stock the product in a dry place, in its original packing (on which the batch number is marked)
- The frequency of inspections to check the equipment is subject to the assessment made by the ortho-prosthetist depending on the patient concerned

### Guarantee

The guarantee is valid for one year as from the date of purchase of the product, if the product has been used in accordance with the manufacturer's instructions.

*N.B. In the event of a complaint, please enclose a return form correctly filled in, incomplete forms cannot be dealt with.*

### EC markings

Groupe G2M hereby declares, in its quality as the manufacturer, that all its parts comply with directive 93/42/EC.