

Restriction on use / Precautions for use

- ⤿ The product must be used within the recommended weight limits and in accordance with the assessment made by an approved ortho-prosthetist depending on the patient concerned
- ⤿ Product maintenance is subject to an assessment made by an approved orthoprosthetist depending on the conditions of use
- ⤿ All assembly and adjustment work must be carried out by qualified personnel

Cleaning / Storage

- ⤿ Stock the product in a dry place, in its original packing (on which the batch number is marked)
- ⤿ The frequency of inspections to check the equipment is subject to the assessment made by the orthoprosthetist depending on the patient concerned


Guarantee

The guarantee is valid for one year as from the date of purchase of the product, if the product has been used in accordance with the manufacturer's instructions.

N.B. In the event of a complaint, please enclose a return form correctly filled in, incomplete forms cannot be dealt with.

EC markings

Groupe G2M hereby declares, in its quality as the manufacturer, that all its parts comply with directive 93/42/EC.



Désignation

Appareil de Perlstein, semelle seule avec tube et dispositif de Perlstein.

Destination

Composant entrant dans la fabrication d'orthèse du membre inférieur avec dispositif de Perlstein.

Pièces détachées

	L42	L43
Longueur de semelle	200 mm	200 mm
Épaisseur de semelle	2 mm	2 mm
Diamètre du tube	8 x 11 mm	6 x 9 mm
Poids	260 g	265 g
Vis pour Perlstein	V0098	V0098
Excentrique	L3502	L3502
Plat inox	L3501	L3501



	L44
Longueur de semelle	150 mm
Épaisseur de semelle	2 mm
Diamètre du tube	6 x 9 mm
Poids	160 g
Vis pour Perlstein	V0099
Excentrique	L3602
Plat inox	L3601



Restriction d'utilisation / Précautions d'emploi

- Les appareils de Perlstein doivent être utilisés suivant l'appréciation de l'orthoprothésiste agréé en fonction du patient à appareiller
- La maintenance du produit reste à l'appréciation de l'orthoprothésiste selon le cadre de l'utilisation
- L'assemblage et l'ajustage doivent être effectués par un personnel qualifié

Entretien / Stockage

- Stocker le produit à l'abri de l'humidité et dans son emballage d'origine (sur lequel est inscrit son numéro de lot)
- La fréquence des visites de contrôle de l'appareil reste à l'appréciation de l'orthoprothésiste en fonction du patient à appareiller

Garantie

La garantie est valable 1 an pour une utilisation conforme aux instructions du fabricant à partir de la date d'achat du produit.

NB : Pour toutes réclamations veuillez joindre un formulaire de retour dûment complété, tout formulaire incomplet ne pourra être traité.

Marquage CE

Groupe G2M déclare, en sa qualité de distributeur, que toutes ses pièces sont conformes à la directive 93/42/CE.



Designation

Perlstein's apparatus, sole only with tube and Perlstein's apparatus.

Purpose

Component used for the manufacture of an orthosis for a lower limb, with Perlstein's apparatus.

Spares parts

	L42	L43
Sole length	200 mm	200 mm
Sole thickness	2 mm	2 mm
Tube diameter	8 x 11 mm	6 x 9 mm
Weight	260 g	265 g
Perlstein screw	V0098	V0098
Eccentric	L3502	L3502
Stainless steel flat	L3501	L3501



	L44
Sole length	150 mm
Sole thickness	2 mm
Tube diameter	6 x 9 mm
Weight	160 g
Perlstein screw	V0099
Eccentric	L3602
Stainless steel flat	L3601

